

www.bioss.com.cn sales@bioss.com.cn techsupport@bioss.com.cn 400-901-9800

# Mouse SOD1 Ready-To-Use IHC Kit

| Cat.No:       | IHC0178M             |
|---------------|----------------------|
| Applications: | ІНС-Р                |
| Reactivity:   | Mouse                |
| Size:         | 50T                  |
| Assay type:   | Immunohistochemistry |
| Sample type:  | FFPE tissue          |
|               |                      |

General Information:

mistry Protocol:

| Number | Component  | Size    | Concentration | Storage                   |
|--------|--|---------|---------------|---------------------------|
| 1      | PBS Buffer (powder)                                  | 2 L × 2 | 20x           | RT                        |
| 2      | Antigen Retrieval Buffer                             | 20 ml   | 100x          | 2-8°C                     |
| 3      | Endogenous Peroxidase Blocking Buffer                | 3 ml    | RTU           | 2-8°C, protect from light |
| 4      | Blocking Buffer                                      | 3 ml    | RTU           | 2-8°C                     |
| 5      | Primary Antibody (Mouse SOD1 Rabbit mAb)             | 6 ml    | RTU           | 2-8°C                     |
| 6      | Secondary Antibody (HRP-Goat anti-Rabbit<br>IgG pAb) | 6 ml    | RTU           | 2-8°C                     |
| 7      | Chromogen Component A                                | 0.3 ml  | RTU           | -20°C,protect from light  |
| 8      | Chromogen Component B                                | 0.3 ml  | RTU           | -20°C                     |
| 9      | Counter Staining Reagent                             | 5 ml    | RTU           | RT                        |
| 10     | Mounting Media                                       | 5 ml    | RTU           | RT                        |
| 11     | Control slide (Mouse brain)                          | 1 slide | RTU           | RT                        |
| 12     | Datasheet  | 1 сору  |               |                           |

Storage andPlease store components at the temperatures indicated on the individual tube labels. TheStability:kit is stable for 6 months from the date of receipt.Immunohistoche

# $1. \ \textbf{Deparaffinization And Rehydration}$

Immerse slides in fresh xylene for 15 minutes and then repeat two more times using separate containers. Immerse slides sequentially in 100%, 95%, 90%, 80%, and 70% ethanol solutions for 5 minutes each. Rinse slides 3 times with distilled water for 5 minutes each.

# 2. Antigen Retrieval

Add  $100 \times$  **Antigen Retrieval Buffer** into distilled water to prepare a  $1 \times$  solution. Boil slides in  $1 \times$  solution at 95°C-100°C for 15 minutes. Move the slides to  $1 \times$  solution at room temperature (RT) and allow them to stand for 20 minutes. Rinse 3 times with **PBS Buffer** (dissolve the powder in 2L distilled water) for 5 minutes each.

#### 3. Block Endogenous Peroxidase

Drain the liquid off the slides and then use a hydrophobic IHC pen to draw circles on the slides around tissue sections. Add 2-4 drops of **Endogenous Peroxidase Blocking Buffer** directly on slides, covering the whole tissue and block slides for 15 minutes at RT. Rinse 3 times with **PBS Buffer** for 5 minutes each.

#### 4. Serum Blocking

Block with 2-4 drops of **Blocking Buffer** for 20 minutes at RT.

#### 5. Primary Antibody Incubation

Drain blocking buffer from slides. Incubate slides with 2-4 drops of **Mouse SOD1 Rabbit mAb** overnight at 4°C or 1-2 hours at RT. Rinse 3 times with **PBS Buffer** for 5 minutes each.

#### 6. Secondary Antibody Incubation

Incubate slides with 2-4 drops of **HRP-Goat anti-Rabbit IgG pAb** for 1-2 hours at RT. Rinse slides 3 times with **PBS Buffer** for 5 minutes each.

#### 7. Signal Development

Remove residual liquid around the tissue section. Add 50ul fresh **DAB Buffer** (**Chromogen Component A : Chromogen Component B : PBS Buffer=1:1:18**) to cover the tissue. Monitor the reaction under the microscope until a brown color is visible (approximate 3-5 minutes at RT). Stop reaction immediately by rinsing with distilled water. Rinse slides 3 times with distilled water for 5 minutes each.

#### 8. Counterstain

Counterstain with an appropriate amount of **Counter Staining Reagent** for 3-5 minutes at RT. Rinse slides with distilled water for 5 minutes. Use 2-4 drops of **Differentiation reagent** to cover the tissue for 30 seconds. Rinse slides twice with distilled water for 5 minutes each.

#### 9. Dehydration Sheet

Immerse slides sequentially in 70%, 80%, 90%, 95%, and 100% ethanol for 5 minutes each at RT. Immerse slides in 2 changes of fresh xylene, 15 minutes each. Drop some **Mounting Media** on the tissue. Mount coverslips.

### Notes:

1. The positive control slide provided in the kit allows you to be sure that the experimental set-up is working properly.

2. Do not allow slides to dry at any time during this procedure.

3. Please don't replace the matching reagents in this product with other manufacturers' products.

4. As DAB is a carcinogen, please take necessary precautions.

5. PBS (reagent 1) can be stored for one week at 4°C after preparation; The antigen retrieval buffer ( $1 \times$  reagent 2) and the chromogenic agent (the mixture of reagents 7 and 8) should be prepared right before each assay.

*Please cite this product as " IHC0178M, Bioss Antibodies". Citation example: " Mouse Tissue sections using SOD1 IHC Kit (IHC0178M, Bioss Antibodies) were stained for SOD1 according to the manufacturer's instructions."* 

Introduction: SOD1 (superoxide dismutase) binds copper and zinc ions and is one of two isozymes responsible for destroying free superoxide radicals in the body. This isozyme is a soluble cytoplasmic protein, acting as a homodimer to convert naturally-occuring but harmful superoxide radicals to molecular oxygen and hydrogen peroxide. The other isozyme is a mitochondrial protein. Mutations in this gene have been implicated as causes of familial amyotrophic lateral sclerosis. Rare transcript variants have been reported for this gene.

# Validation Data



Immunohistochemical analysis of paraffin embedded mouse colon tissue slide using IHC0178M (Mouse SOD1 IHC Kit).